

District	Beam Breaker ID #	Date Calibrated	400=3200	500=4000	600=4800	650=5200	700=5600	Expected Beam Breaker Reading in psi					By	Remarks
			Actual Compression Machine Screen Readings in Pounds					400	500	600	650	700		
6	301659	2/3/2020	3250	4030	4810	5180	5550	2%	1%	0%	0%	-1%	PTK/MB	
6	301659	2/7/2022	3300	4090	4820	5230	5600	3%	2%	0%	1%	0%	MB	
6	301659	2/6/2023	3260	4035	4800	5205	5565	2%	1%	0%	0%	-1%	MB	
6	301659	1/31/2025	3245	4015	4800	5180	5560	1%	0%	0%	0%	-1%	AB	
6	301659	1/29/2026	3300	4095	4810	5200	5580	3%	2%	0%	0%	0%	JS	